

Request for Quotation

REF: AE/AK/PWC/05

Sub : Quotation for supply of: **Online Air Heater**

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover in order to reach us before 5/5/2012 in the form of hard copy / soft copy.

The soft copy of the model quotation can also be downloaded from our website

Technical details of the store are as follows;

Air Heater

Rating – 200 KW/440 VOLTS, 3 PHASE.

Air Flow Rate: 0.6 kg/s

Max. temp. 500 deg c. at a pressure of 20 bar

Inner Chamber made out of MS of 5mm thick.

Insulation: 150mm Insulation with Ceramic Fibre Blanket of 64 density.

Inlet & Out let 4" pipe made out of SS with necessary flanges.

Control Panel: A control panel consisting of Digital PID Temp. Controllers (02 nos).

Thyristor Power Pack, On/Off switches, Indication lamps, fuse carriers, Main switches blower starter & other Electrical accessories

Best regards,
Sincerely,

A. Kushari
Department of Aerospace Engineering
IIT Kanpur, India.