



Indian Institute of Technology
Department of Biological Sciences and Bioengineering
KANPUR – 208 016, INDIA

Dr. Dharendra S. Katti
Professor

Tel: 91-512-259-4028
Fax: 91-512-259-4010
Email: dsk@iitk.ac.in

Dated: 01/08/2013
Closing Date: 08/08/13

Enquiry No.: DSK/BSBE/13-14/NC-02

Request for submission of quotation for "Diode Laser"

Sealed quotation(s) are required on or before August 08 2013 latest by 17.00 hrs with all technical specifications for the supply of items;

(Technical Specification Required)

1. 532 diode LASER.
2. Should be automated to move in a semi-circular way, for length of around 3-10mm (3-5mm is desirable).
3. The spot size needed is 50 μ m to 100 μ m.
4. Power range 100-600mW.
5. Time for diode application is 0.1 sec. Time setup should be automated.

Note:

1. The envelope must be inscribed with word "Quotations against enquiry no.13-14/NC-02"
2. All quotations must reach by **August 08, 2013**, at or before 17.00 hrs.
3. Please provide separate lines for the initial price and discounted price with minimum 90 days validity for the quotation (120 days would be desirable)
4. The delivery period should not be more than 4 weeks.
5. Send complete details of the products.
6. **Minimum 2 years of warranty**

Dhirendra S. Katti