

Inquiry no.: IITK/CE/AS/06-2015/02 Closing date and time: 17.06.2015 at 5 pm Date: 03.06.2015

Sub: Call for quotation for supply and installation of a multi parameter pH meter

Sealed quotations (**Technical and Financial bids separately**) are invited from authorized suppliers for items and their specifications given below before 5 pm of 17.06.2015.

The quotation for supply and installation of **a multi parameter pH meter** should be sent in two parts in sealed envelopes, clearly marked as "**Technical Bid**" and "**Financial Bid**". The Technical Bid should contain detailed technical specifications of the product being offered and **should not mention any prices**. The Financial Bid should include the detailed price quotation clearly, including the cost of the equipment, taxes, service charges, shipping and handling charges, if any. Our organization is an educational institute of repute and liable to get maximum <u>education discount</u> from manufacturer. Please specify it, separately.

| S. | Product Name & Specifications | Quantity |
|-----|---|----------|
| No. | | • |
| 1. | Multi parameter pH meter and electrode stand | 1 |
| | The pH meter should be a microprocessor controlled, freely programmable and | |
| | user friendly system with ease of operation and should be able to analyse | |
| | multiple parameters in environmental samples from diverse matrices such as the ones listed below: | |
| | pH: | |
| | • Range: 0.0 to 14.0 pH or wider range | |
| | • Resolution: ±0.001 pH or better (user selectable) | |
| | • Accuracy: ±0.005 pH or better | |
| | • Calibration with 3 points or higher | |
| | • Temperature compensation: Automatic, 0.0 to 100.0 °C | |
| | ORP: | |
| | • Range: 1 to ±1999.9mV or higher | |
| | • Resolution: 0.1 mV or better | |
| | • Accuracy: ±0.2 mV or better | |
| | • Calibration with at least two points (specify how many) | |
| | • Temperature compensation: Automatic, 0.0 to 100.0 °C | |
| | Ions: | |



| • | Fluoride and other ions | | | | |
|--|---|--|--|--|--|
| • | Range: 0.001 ppm to 14000 ppm or wider | | | | |
| | Resolution: 3 digits or better (user selectable) | | | | |
| • | Accuracy: 0.5 % Full scale (monovalent) / 1 % Full scale (divalent) | | | | |
| • | 5 or more calibration points | | | | |
| • | 5 of more canoration points | | | | |
| Condu | Conductivity: | | | | |
| | Range: Should span 0 to 2000 mS/cm or higher with multiple ranges. | | | | |
| | Resolution: 0.05% of full scale or better | | | | |
| | Accuracy: ± 1 % of Full scale ± 1 LSD or better | | | | |
| | Calibration point: Multiple points with each range spanned with | | | | |
| • | minimum 1 point | | | | |
| • | 1 | | | | |
| | Temperature compensation: 0 to 50° C or better, automatic | | | | |
| • | Adjustable cell constant | | | | |
| TDC | | | | | |
| TDS: | | | | | |
| | Range: 0 to 100 g/L or wider | | | | |
| | Resolution: 0.05 % of full scale or better | | | | |
| • | Accuracy: ± 1 % Full scale ± 1 LSD | | | | |
| ٠ | TDS conversion factor: 0.4 to 1.0 | | | | |
| | | | | | |
| Salinity: | | | | | |
| • | Range: 0.0 to 50.0 ppt or higher | | | | |
| • | Resolution: 0.1 ppt or better | | | | |
| • | Accuracy: ± 1 % Full scale ± 1 LSD or better | | | | |
| | - | | | | |
| Dissolved Oxygen: | | | | | |
| | Range: 0 to 40 mg/L or higher | | | | |
| | Resolution: 0.1 mg/L or better | | | | |
| | Accuracy: $\pm 0.2 \text{ mg/L}$ % Full scale $\pm 1 \text{ LSD}$ or better | | | | |
| • | Calibration points: preferred, at least one | | | | |
| | Range for % saturation of oxygen: 0 to 600 | | | | |
| • | | | | | |
| • | Resolution for % saturation of oxygen: 1% or better | | | | |
| • | Accuracy for % saturation of oxygen: $\pm 2\%$ or better | | | | |
| Temperature: | | | | | |
| - | Range: 0.0 to 100.0 °C or higher | | | | |
| | Resolution: 0.1 °C | | | | |
| • | Accuracy: $\pm 0.3^{\circ}$ C or better | | | | |
| - | | | | | |
| Memory: Minimum 100 data sets | | | | | |
| Output: USB / IrDA Interface Adapter / RS232 / RS232C | | | | | |
| Power supply: Long life compatible batteries, 9 V DC adapter, AC 230-240V | | | | | |
| 50/60 Hz | | | | | |
| | | | | | |

Operating Temperature: 5-45 °C (ambient)



Environmental Geochemistry Laboratory Department of Civil Engineering Indian Institute of Technology Kanpur

| | Operating Conditions: 5-85 % (non-condensing) | |
|----|---|--------|
| | Weight: Preference for lower weight Diagnostic message display should be present. | |
| 2. | Compatible electrodes for parameters listed above (quote electrodes with model numbers and clearly specify whether these are included in the price of item 1):-a) Stainless steel ATC probeb) Glass pH electrode, refillablec) Fluoride electroded) ORP electrodee) TDS, Salinity, Conductivity probesf) Galvanic DO probe with integral temperature sensor | 2 sets |
| 3. | NIST traceable standards for calibrating above parameters. Quote separately their make, volumes (mL) and expiry dates and clearly specify whether these are included in the price of item 1. | 1 set |

Terms and Conditions:

- The vendor should supply list of installation (minimum 5, in last two year) in India of the same model quoted against this enquiry preferably at IITK.
- Manufacturer should have appropriate certification.
- If the Financial Bid is included in the Technical Bid, then the quotation will be rejected.
- Quotation should have minimum validity of 60 days from the date of opening.
- Delivery period should be within 60 days from the receipt of the purchase order. Shorter delivery time may be given preference.
- Taxes, packaging, forwarding freight charges, if any should be mentioned.
- Quotation should carry proprietary certificates and authorization letters/certificates.
- Prices should include installation and training of the equipment.
- Provide contact number/address for complaint, else quotes may be rejected.
- The warranty period for the meter should be 3 years and at least 6 months for the electrodes from the date of installation.
- The firms may also quote for optional accessories which will extend the capability or ease of use of the equipment.



- All quotations should be in the currency of the country of origin of the instrument and FOB and CIF, Delhi (if imported), and also converted to ₹.
- The Institute is exempted from excise duty and pays a nominal customs duty of ~5% under Govt. of India notifications 10/97 and 51196, respectively. Custom Duty exemption certificate under notification 51196 and road permit will be provided if applicable.
- Normal payment terms for the Institute will be applicable (90% on delivery of the items and the remaining 10% after satisfactory installation/ inspection).
- The Institute reserves the right for accepting and rejecting any quotations without assigning any reason thereof. Also, the Institute reserves the right to reject or accept all or any of the offers made above.

Thanking you,

Sincerely, Dr. Abhas Singh Assistant Professor, FB-306, Department of Civil Engineering, I.I.T. Kanpur, Kanpur- 208016, Uttar Pradesh, India.